

a measurable difference...



Tuft Withdrawal Fixtures

Model: C0051-A


The IDM Tuft Withdrawal Fixtures are used alongside a tensile machine to measure the force required to withdraw a single tuft or loop of pile from a carpet – that is, the binding force between carpet pile and backing.

The carpet sample is held down by a steel plate, and a pair of surgical forceps is clamped to one end of the tuft to be tested, or a hook threaded through one loop.

The forceps or hook are linked to a digital force gauge, which is raised at a steady rate by a small electronic motor. Tension on the tuft or loop is thus increased and the digital force gauge indicates the maximum force needed to withdraw it.

Applications:

- Textile floor covering which has a tuft (cut) loop pile yarn structure
- Woven: individual loops
- Woven: individual cut tufts



Loop Withdrawal Fixture



Tuft Withdrawal Fixture

10-11 Colrado Court, Hallam, Victoria 3803 Australia.

Phone: +61 3 9708 6885

Fax: +61 3 9708 6770

Website: www.idminstruments.com.au

Email: ids@idminstruments.com.au

All sizes and dimensions shown on this data sheet are averages only and images are as accurate as the medium allows. Products and details shown may change without notice as IDM Instruments Pty Ltd[®] reserves the right to make alterations to its product range from time to time.

Copyright © 2006 IDM Instruments Pty Ltd[®]. All Rights Reserved.